



ARJOWIGGINS

TECHNICAL INFORMATION NOTICE
TO Nan Tong Fuhua Medical Packing Co.,LTD

Address : 1#A Area XinKai industrial Park,
XinFu Road ,
Nan Tong Development Zone,
Jiangsu Province
CHINA

The representative product, of the range in which the following product belongs, has been evaluated as "sufficiently germ-proof" for the microbiological analysis "germ proofness" in moisture (subclause 2.14) and with passage of air (subclause 2.15) according to DIN 58953 part 6.

PRODUCT DESIGNATION

ETHYPEL DSR70 60GSM

CODE

0A24000

This information is based on our current knowledge at the date hereof and on the results given by ISEGA which has performed the test (Report Nber 1494/148a of the 30/03/2005). It applies to the sole product individually referred to herein.

Friday, 16 May 2008
Regulatory Affairs Coordinator,

JM GAILLARD