



ARJOWIGGINS

**TECHNICAL INFORMATION NOTICE**  
**TO Nan Tong Fuhua Medical Packing Co.,LTD**

**Address :** 1#A Area XinKai industrial Park,  
XinFu Road ,  
Nan Tong Development Zone,  
Jiangsu Province  
CHINA

The porous product hereafter :

**PRODUCT DESIGNATION**

**PROPYPEL PF 70GSM**

**CODE**

**0779000**

complies with the EN ISO 11607-Part 1 requirements : *Packaging for terminally sterilized medical devices - Part 1 requirement for materials, sterile barrier systems and packaging systems*, when :

- used according following sterilization method :
  - o Steam (134°C / 23 min)
  - o Ethylene oxyde (54°C / 90min)
  - o Gamma ray (25kGy)
- within 5 years after its manufacturing date,
- with stockage according to the following conditions
  - o Stored in the original packaging (protective polyethylene wrapper)
  - o Stored in a cool, dry location away from strong sunlight or UV light (30 to 60% relative humidity; 50 to 104°F / 10 to 40°C)
  - o Use the earliest materials received first

for a compatibility with respect to forming and sealing evaluated trough lab sealing bar tester using PP film.

This information is based on our current knowledge. It applies to the sole product individually referred to herein.

Friday, 16 May 2008  
Regulatory Affairs Coordinator,

**JM GAILLARD**